



BRIGON CO₂ Indicator TESTORYT



CO₂-Indicator is available in 3 different designs:

- CO₂ indicator with measuring range 0 ... 10 Vol% CO₂
- CO₂ indicator with measuring range 0 ... 20 Vol% CO₂
- CO₂ indicator with measuring range 0 ... 60 Vol% CO₂

A flue gas sample is taken by the aspirator assembly into the indicator flask. Absorption of CO₂ from the flue gas sample leads to the volume contraction of the liquid and causes the diaphragm at the bottom to stretch upwards and push the liquid to the level indicating the CO₂ value in Vol%.

Delivery content

- CO₂ indicator 0 ... 10 Vol% or 0 ... 20 Vol% or 0 ... 60 Vol%
- Aspirator assembly
- Refill liquid, red, 70 ml bottle

Item no. 4110	CO₂ Indicator TESTORYT 0 ... 10 Vol% complete
Item no. 4115	CO₂ Indicator TESTORYT 0 ... 10 Vol% without liquid, cpl.

Item no. 4120	CO₂ Indicator TESTORYT 0 ... 20 Vol% complete
Item no. 4125	CO₂ Indicator TESTORYT 0 ... 20 Vol% without liquid, cpl.

Item no. 4160	CO₂ Indicator TESTORYT 0 ... 60 Vol% complete
Item no. 4165	CO₂ Indicator TESTORYT 0 ... 60 Vol% without liquid, cpl.